

Certificado No. (Certificate #):		64500		Equipo (Instrument):		Medidor de Espesores	
Fecha de Calibración (Calibration Date):		2025-12-17		Marca (Brand):		Mitutoyo	
Patrón (Standard)		UBP (UUT)	Error (Error)	E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
1.00 mm		1.00 mm	0 µm	± 25.4 µm		±	5.8E+00 µm
1.50 mm		1.50 mm	0 µm	± 25.4 µm		±	5.8E+00 µm
2.00 mm		2.00 mm	0 µm	± 25.4 µm		±	5.8E+00 µm
3.00 mm		3.00 mm	0 µm	± 25.4 µm		±	5.8E+00 µm
4.00 mm		4.00 mm	0 µm	± 25.4 µm		±	5.8E+00 µm
5.00 mm		5.00 mm	0 µm	± 25.4 µm		±	5.8E+00 µm
7.00 mm		6.99 mm	-10 µm	± 25.4 µm		±	5.8E+00 µm
10.00 mm		9.99 mm	-10 µm	± 25.4 µm		±	5.8E+00 µm
12.00 mm		11.99 mm	-10 µm	± 25.4 µm		±	5.8E+00 µm