

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.775 %HR	27.9 %HR	-2.875 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.509 %HR	49.5 %HR	-1.009 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	71.4 %HR	0.748 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	90.519 %HR	93.9 %HR	3.381 %HR	± 5 %HR	± 1.9E+00 %HR

