

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.775 %HR	28.5 %HR	-2.275 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.509 %HR	50.1 %HR	-0.409 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	71.5 %HR	0.848 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	90.519 %HR	90.4 %HR	-0.119 %HR	± 5 %HR	± 1.9E+00 %HR

