

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-24.889 °C	-25.3 °C	-0.411 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.020 °C	0.3 °C	0.320 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.294 °C	5.0 °C	-0.294 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.408 °C	10.1 °C	-0.308 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.4 °C	-0.050 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.3 °C	-0.163 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.090 °C	39.8 °C	-0.290 °C	± 0.45 °C	± 4.0E-01 °C

