

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -24.889 °C        | -24.5 °C  | 0.389 °C      | ± 0.95 °C            | ± 4.0E-01 °C                |
| 0 °C              | -0.020 °C         | 0.1 °C    | 0.120 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 5 °C              | 5.294 °C          | 5.3 °C    | 0.006 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 10 °C             | 10.408 °C         | 10.3 °C   | -0.108 °C     | ± 0.45 °C            | ± 4.0E-01 °C                |
| 25 °C             | 25.450 °C         | 25.3 °C   | -0.150 °C     | ± 0.45 °C            | ± 4.0E-01 °C                |
| 30 °C             | 30.463 °C         | 30.1 °C   | -0.363 °C     | ± 0.45 °C            | ± 4.0E-01 °C                |
| 40 °C             | 40.090 °C         | 40.5 °C   | 0.410 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |

