

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
4 °C	4.021 °C	4.5 °C	0.479 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.272 °C	15.6 °C	0.328 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.782 °C	30.1 °C	0.318 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.971 %HR	34.6 %HR	2.629 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.803 %HR	46.9 %HR	-2.903 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	62.513 %HR	61.6 %HR	-0.913 %HR	± 3.0 %HR	± 1.9E+00 %HR