

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 5	4.99971	0.004	± 0.1	± 3.5E-01 in					
0 a 10	10.00025	-0.003	± 2.2	± 3.5E-01 in					
0 a 15	14.99959	0.005	± 3.2	± 3.5E-01 in					
0 a 20	19.99961	0.005	± 4.2	± 3.5E-01 in					
0 a 25	25.00037	-0.004	± 5.2	± 3.5E-01 in					
0 a 30	29.99936	0.008	± 6.2	± 3.5E-01 in					
0 a 35	35.00015	-0.002	± 7.2	± 3.5E-01 in					
0 a 40	40.00001	0.000	± 8.2	± 3.5E-01 in					
0 a 45	45.00029	-0.004	± 9.2	± 3.5E-01 in					
0 a 50	49.99955	0.005	± 10.2	± 3.5E-01 in					