

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
5 °C	4.993 °C	5.0 °C	-0.043 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.7 °C	0.214 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.090 °C	40.2 °C	0.100 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.713 %HR	47.0 %HR	-3.713 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.366 %HR	78.0 %HR	-2.366 %HR	± 5.0 %HR	± 1.9E+00 %HR