

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.772 °C	29.9 °C	0.128 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	31.2 %HR	1.867 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	50.2 %HR	0.182 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.675 %HR	74.5 %HR	-2.175 %HR	± 3.0 %HR	± 1.9E+00 %HR