

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.772 °C	30.0 °C	0.228 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	27.9 %HR	-1.433 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	48.5 %HR	-1.518 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.675 %HR	76.0 %HR	-0.675 %HR	± 3.0 %HR	± 1.9E+00 %HR