

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.772 °C	29.7 °C	-0.072 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	28.3 %HR	-1.033 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	48.7 %HR	-1.318 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.675 %HR	76.2 %HR	-0.475 %HR	± 3.0 %HR	± 1.9E+00 %HR