

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.714 °C	30.5 °C	-0.214 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.149 %HR	27.0 %HR	-2.149 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.578 %HR	47.9 %HR	-1.678 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.773 %HR	76.2 %HR	1.427 %HR	± 3.0 %HR	± 1.9E+00 %HR