

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.714 °C	30.7 °C	-0.014 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.149 %HR	26.6 %HR	-2.549 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.578 %HR	47.7 %HR	-1.878 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.773 %HR	77.9 %HR	3.127 %HR	± 3.0 %HR	± 1.9E+00 %HR