

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.714 °C	30.7 °C	-0.014 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.149 %HR	29.7 %HR	0.551 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.578 %HR	50.3 %HR	0.722 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.773 %HR	77.5 %HR	2.727 %HR	± 3.0 %HR	± 1.9E+00 %HR