

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
42 %HR	41.818 %HR	40.37 %HR	-1.448 %HR	± 5.0 %HR	± 1.9E+00 %HR