

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
26 %HR	26.480 %HR	26.04 %HR	-0.440 %HR	± 5.0 %HR	± 1.9E+00 %HR