

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
28 %HR	28.525 %HR	29.70 %HR	1.175 %HR	± 5.0 %HR	± 1.9E+00 %HR