

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.0 °C	-0.062 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.3 %HR	-0.270 %HR	± 3 %HR	± 7.4E-02 %HR
50 %HR	50.100 %HR	50.2 %HR	0.100 %HR	± 3 %HR	± 7.4E-02 %HR
70 %HR	70.448 %HR	70.9 %HR	0.452 %HR	± 3 %HR	± 7.4E-02 %HR