

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.100 °C	0.5 °C	0.400 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.245 °C	20.4 °C	0.155 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.588 °C	35.5 °C	-0.088 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.026 %HR	33.4 %HR	4.374 %HR	± 5.0 %HR	± 1.9E+00 %HR
55 %HR	54.834 %HR	52.5 %HR	-2.334 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	79.783 %HR	75.4 %HR	-4.383 %HR	± 5.0 %HR	± 1.9E+00 %HR