

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.100 °C	0.5 °C	0.400 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.245 °C	20.5 °C	0.255 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.588 °C	35.4 °C	-0.188 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.026 %HR	33.4 %HR	4.374 %HR	± 5.0 %HR	± 1.9E+00 %HR
55 %HR	54.834 %HR	52.7 %HR	-2.134 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	79.783 %HR	75.8 %HR	-3.983 %HR	± 5.0 %HR	± 1.9E+00 %HR