

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | 0.100 °C                   | 0.6 °C             | 0.500 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.245 °C                  | 20.5 °C            | 0.255 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 35 °C                      | 35.588 °C                  | 35.3 °C            | -0.288 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.026 %HR                 | 34.0 %HR           | 4.974 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 55 %HR                     | 54.834 %HR                 | 53.6 %HR           | -1.234 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 80 %HR                     | 79.783 %HR                 | 74.8 %HR           | -4.983 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |