

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.452 °C	0.0 °C	0.452 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.205 °C	20.2 °C	-0.005 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.588 °C	35.3 °C	-0.288 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.355 %HR	33.0 %HR	2.645 %HR	± 5.0 %HR	± 1.9E+00 %HR
55 %HR	53.392 %HR	52.9 %HR	-0.492 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.141 %HR	77.8 %HR	-2.341 %HR	± 5.0 %HR	± 1.9E+00 %HR