

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.714 °C	29.9 °C	-0.814 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.149 %HR	24 %HR	-5.149 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.578 %HR	45 %HR	-4.578 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.773 %HR	76 %HR	1.227 %HR	± 5.0 %HR	± 1.9E+00 %HR