

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.772 °C	29.7 °C	-0.072 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	28.4 %HR	-0.933 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	51.5 %HR	1.482 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	77.084 %HR	75.9 %HR	-1.184 %HR	± 3.0 %HR	± 1.9E+00 %HR