

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.772 °C	30.0 °C	0.228 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	29.8 %HR	0.467 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	49.9 %HR	-0.118 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	77.084 %HR	74.2 %HR	-2.884 %HR	± 3.0 %HR	± 1.9E+00 %HR