

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60 RPM	60.0 RPM	0.0 RPM	± 0.11 RPM	±	5.8E-02 RPM
600 RPM	600.0 RPM	0.0 RPM	± 0.16 RPM	±	5.8E-02 RPM
6000 RPM	6000.0 RPM	0.0 RPM	± 0.7.0 RPM	±	5.8E-02 RPM
60000 RPM	60000 RPM	0.0 RPM	± 7.0 RPM	±	8.0E-01 RPM