

Patrón <i>(Standard)</i>		UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
7 Hz	0.400 g-pk	0.39 g-pk	-0.010 g-pk	± 0.03 g-pk	±	1.1E-02 g-pk
10 Hz	0.800 g-pk	0.79 g-pk	-0.010 g-pk	± 0.03 g-pk	±	1.7E-02 g-pk
30 Hz	1.000 g-pk	0.99 g-pk	-0.010 g-pk	± 0.03 g-pk	±	2.1E-02 g-pk
50 Hz	1.000 g-pk	0.99 g-pk	-0.010 g-pk	± 0.03 g-pk	±	2.1E-02 g-pk
100 Hz	1.000 g-pk	0.99 g-pk	-0.010 g-pk	± 0.03 g-pk	±	2.1E-02 g-pk
500 Hz	1.000 g-pk	0.99 g-pk	-0.010 g-pk	± 0.03 g-pk	±	2.1E-02 g-pk