

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.100 °C	0.3 °C	0.200 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.124 °C	5.6 °C	0.476 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.218 °C	10.7 °C	0.482 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	54.609 %HR	60.3 %HR	5.691 %HR	± 3.0 %HR	± 1.9E+00 %HR