

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.481 °C	22.0 °C	-0.481 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.8 °C	-0.249 °C	± 1.0 °C	± 4.0E-01 °C
31 °C	30.965 °C	30.6 °C	-0.365 °C	± 1.0 °C	± 4.0E-01 °C
48 %HR	47.953 %HR	46.3 %HR	-1.653 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	57.462 %HR	53.5 %HR	-3.962 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	65.1 %HR	-2.280 %HR	± 5.0 %HR	± 1.9E+00 %HR