

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.481 °C	22.8 °C	0.319 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.6 °C	0.551 °C	± 1.0 °C	± 4.0E-01 °C
31 °C	30.965 °C	31.6 °C	0.635 °C	± 1.0 °C	± 4.0E-01 °C
48 %HR	47.953 %HR	50.3 %HR	2.347 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	57.462 %HR	60.2 %HR	2.738 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	71.4 %HR	4.020 %HR	± 5.0 %HR	± 1.9E+00 %HR