

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.481 °C	23.1 °C	0.619 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.8 °C	0.751 °C	± 1.0 °C	± 4.0E-01 °C
31 °C	30.965 °C	31.5 °C	0.535 °C	± 1.0 °C	± 4.0E-01 °C
48 %HR	47.953 %HR	51.4 %HR	3.447 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	57.462 %HR	54.6 %HR	-2.862 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	65.8 %HR	-1.580 %HR	± 5.0 %HR	± 1.9E+00 %HR