

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.481 °C	22.6 °C	0.119 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.3 °C	0.251 °C	± 1.0 °C	± 4.0E-01 °C
31 °C	30.965 °C	31.3 °C	0.335 °C	± 1.0 °C	± 4.0E-01 °C
48 %HR	47.953 %HR	50.2 %HR	2.247 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	57.462 %HR	59.7 %HR	2.238 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	66.9 %HR	-0.480 %HR	± 5.0 %HR	± 1.9E+00 %HR