

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.481 °C	22.3 °C	-0.181 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.4 °C	0.351 °C	± 1.0 °C	± 4.0E-01 °C
31 °C	30.965 °C	31.3 °C	0.335 °C	± 1.0 °C	± 4.0E-01 °C
48 %HR	47.953 %HR	51.2 %HR	3.247 %HR	± 5.0 %HR	± 1.9E+00 %HR
57 %HR	57.462 %HR	60.3 %HR	2.838 %HR	± 5.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	65.3 %HR	-2.080 %HR	± 5.0 %HR	± 1.9E+00 %HR