

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
17 °C	16.696 °C	16.7 °C	0.004 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	21.048 °C	21.1 °C	0.052 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	22.161 °C	22.3 °C	0.139 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	27.716 °C	27.8 °C	0.084 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.224 %HR	31.9 %HR	0.676 %HR	± 5.0 %HR	± 1.9E+00 %HR
40 %HR	42.339 %HR	42.8 %HR	0.461 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	61.347 %HR	62.1 %HR	0.753 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	78.116 %HR	79.0 %HR	0.884 %HR	± 5.0 %HR	± 1.9E+00 %HR