

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
17 °C	16.696 °C	16.8 °C	0.104 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	21.048 °C	21.1 °C	0.052 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	22.161 °C	22.2 °C	0.039 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	27.716 °C	27.8 °C	0.084 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.224 %HR	32.4 %HR	1.176 %HR	± 5.0 %HR	± 1.9E+00 %HR
40 %HR	42.339 %HR	43.0 %HR	0.661 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	61.347 %HR	62.4 %HR	1.053 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	78.116 %HR	79.5 %HR	1.384 %HR	± 5.0 %HR	± 1.9E+00 %HR