

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.342 °C	15.6 °C	0.258 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.520 °C	25.8 °C	0.280 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.604 °C	30.8 °C	0.196 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.936 %HR	64.5 %HR	4.564 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.670 %HR	67.8 %HR	3.130 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	72.9 %HR	2.534 %HR	± 3.0 %HR	± 1.9E+00 %HR