

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.424 °C	-29.6 °C	0.824 °C	± 1.0 °C	± 4.0E-01 °C
-25 °C	-25.531 °C	-24.8 °C	0.731 °C	± 1.0 °C	± 4.0E-01 °C
-15 °C	-15.573 °C	-14.9 °C	0.673 °C	± 1.0 °C	± 4.0E-01 °C
-10 °C	-10.830 °C	-10.2 °C	0.630 °C	± 1.0 °C	± 4.0E-01 °C
2 °C	2.306 °C	2.9 °C	0.594 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.453 °C	8.9 °C	0.447 °C	± 0.5 °C	± 4.0E-01 °C