

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.424 °C	-29.8 °C	0.624 °C	± -1.1 °C	± 4.0E-01 °C
-25 °C	-25.531 °C	-24.9 °C	0.631 °C	± -1.1 °C	± 4.0E-01 °C
-15 °C	-15.573 °C	-15.1 °C	0.473 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.830 °C	-10.4 °C	0.430 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	2.306 °C	2.4 °C	0.094 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.453 °C	8.4 °C	-0.053 °C	± 0.5 °C	± 4.0E-01 °C