

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
2,004.8 lx	2,001.1 lx	3.7 lx	± 65 lx	±	2.9E+00 lx
1,077.0 lx	1,074.0 lx	3.0 lx	± 37 lx	±	1.1E+00 lx
500.0 lx	498.0 lx	2.0 lx	± 20 lx	±	2.9E-01 lx
21.5 lx	20.0 lx	1.5 lx	± 6 lx	±	2.9E-02 lx
10.6 lx	10.1 lx	0.5 lx	± 5 lx	±	2.9E-02 lx