

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09963	0.370	± 0.6	±	4.3E-01 mm						
0 a 0.5	0.49931	0.690	± 0.8	±	4.3E-01 mm						
0 a 1	0.99960	0.400	± 1	±	4.3E-01 mm						
0 a 2	1.99959	0.410	± 1.4	±	4.8E-01 mm						
0 a 3	2.99962	0.380	± 1.8	±	4.8E-01 mm						
0 a 4	3.99963	0.370	± 2.2	±	4.8E-01 mm						
0 a 5	4.99966	0.340	± 2.6	±	4.8E-01 mm						
0 a 6	6.00031	-0.310	± 3	±	3.6E-01 mm						
0 a 7	6.99975	0.250	± 3.4	±	3.6E-01 mm						
0 a 8	7.99980	0.200	± 3.8	±	3.6E-01 mm						