

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	34.555 °C	34.6 °C	0.045 °C	± 0.4 °C	± 4.0E-01 °C
30 %HR	29.783 %HR	31.0 %HR	1.217 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.241 %HR	50.8 %HR	1.559 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	67.789 %HR	69.4 %HR	1.611 %HR	± 3.0 %HR	± 1.9E+00 %HR