

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
30 °C	34.555 °C	34.5 °C	-0.055 °C	± 0.2 °C	± 4.0E-01 °C
30 %HR	29.783 %HR	31.1 %HR	1.317 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.241 %HR	51.0 %HR	1.759 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	67.789 %HR	69.1 %HR	1.311 %HR	± 3.0 %HR	± 1.9E+00 %HR