

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 34.555 °C                  | 34.6 °C            | 0.045 °C               | ± 0.2 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.783 %HR                 | 31.0 %HR           | 1.217 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 49.241 %HR                 | 50.9 %HR           | 1.659 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 67.789 %HR                 | 69.2 %HR           | 1.411 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |