

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	34.555 °C	34.5 °C	-0.055 °C	± 0.2 °C	± 4.0E-01 °C
30 %HR	29.783 %HR	31.1 %HR	1.317 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.241 %HR	50.9 %HR	1.659 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	67.789 %HR	69.6 %HR	1.811 %HR	± 3.0 %HR	± 1.9E+00 %HR