

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.971 °C	15.6 °C	0.629 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.3 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	32.1 °C	0.032 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	19.3 %HR	-1.045 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.100 %HR	50.3 %HR	0.200 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.499 %HR	70.1 %HR	-0.399 %HR	± 5.0 %HR	± 1.9E+00 %HR