

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04940	0.600		± 4.3E-01 mm					
0 a 0.1	0.09966	0.340		± 4.3E-01 mm					
0 a 0.15	0.15001	-0.010		± 4.3E-01 mm					
0 a 1	0.99914	0.860		± 4.3E-01 mm					
0 a 2	1.99911	0.890		± 4.8E-01 mm					
0 a 3	2.99917	0.830		± 4.8E-01 mm					
0 a 4	3.99927	0.730		± 4.8E-01 mm					
0 a 5	4.99930	0.700		± 4.8E-01 mm					
0 a 6	5.99932	0.680		± 3.6E-01 mm					
0 a 7	6.99939	0.610		± 3.6E-01 mm					
0 a 8	7.99945	0.550		± 3.6E-01 mm					
0 a 9	8.99947	0.530		± 3.6E-01 mm					
0 a 10	9.99950	0.500		± 3.6E-01 mm					
0 a 11	10.99952	0.480		± 3.6E-01 mm					
0 a 12	11.99954	0.460		± 3.6E-01 mm					
0 a 13	12.99957	0.430		± 3.6E-01 mm					
0 a 14	13.99960	0.400		± 3.6E-01 mm					
0 a 15	14.99962	0.380		± 3.6E-01 mm					
0 a 16	15.99964	0.360		± 3.6E-01 mm					
0 a 17	16.99965	0.350		± 3.6E-01 mm					
0 a 18	17.99968	0.320		± 3.6E-01 mm					
0 a 19	18.99974	0.260		± 3.6E-01 mm					
0 a 20	19.99980	0.200		± 3.6E-01 mm					
0 a 21	20.99985	0.150		± 5.5E-01 mm					
0 a 22	21.99990	0.100		± 5.5E-01 mm					
0 a 23	22.99992	0.080		± 5.5E-01 mm					
0 a 24	24.00000	0.000		± 5.5E-01 mm					
0 a 25	25.00005	-0.050		± 5.5E-01 mm					
0 a 26	26.00011	-0.110		± 5.5E-01 mm					
0 a 27	27.00013	-0.130		± 5.5E-01 mm					
0 a 28	28.00020	-0.200		± 5.5E-01 mm					
0 a 29	29.00026	-0.260		± 5.5E-01 mm					
0 a 30	30.00027	-0.270		± 5.5E-01 mm					