

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 1.00 mm                  | 1.00 mm          | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 1.25 mm                  | 1.25 mm          | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 1.50 mm                  | 1.50 mm          | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 2.00 mm                  | 2.00 mm          | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 3.00 mm                  | 3.00 mm          | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 5.00 mm                  | 5.00 mm          | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 10.00 mm                 | 10.00 mm         | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 20.00 mm                 | 20.00 mm         | 0 μm                 | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |
| 25.00 mm                 | 24.99 mm         | -10 μm               | ± 0.03 μm                   | ±                                  | 5.8E+00 μm |