

Certificado No. (Certificate #):		64918		Equipo (Instrument):		Inclinometro	
Fecha de Calibración (Calibration Date):		2026-01-14		Marca (Brand):		KLEIN TOOLS	
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)		Incertidumbre (Uncertainty)	
0.50 °		0.4 °	-0.1 °	± 0.2 °		±	6.40E-02 °
3.00 °		3.0 °	0.0 °	± 0.2 °		±	6.40E-02 °
5.00 °		5.0 °	0.0 °	± 0.2 °		±	6.40E-02 °
15.00 °		15.1 °	0.1 °	± 0.2 °		±	6.40E-02 °
25.00 °		25.0 °	0.0 °	± 0.2 °		±	6.40E-02 °
30.00 °		30.0 °	0.0 °	± 0.2 °		±	7.60E-02 °
45.00 °		45.0 °	0.0 °	± 0.2 °		±	7.60E-02 °
60.00 °		60.0 °	0.0 °	± 0.2 °		±	7.60E-02 °
90.00 °		90.0 °	0.0 °	± 0.2 °		±	7.60E-02 °