

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	3.910 °C	3.7 °C	-0.210 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.131 °C	15.1 °C	-0.031 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.872 °C	29.9 °C	0.028 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.325 %HR	30.1 %HR	-0.225 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.261 %HR	48.2 %HR	-1.061 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.486 %HR	61.5 %HR	-2.986 %HR	± 3.0 %HR	± 1.9E+00 %HR