

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
4 °C	4.111 °C	4.1 °C	-0.011 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.881 °C	15.0 °C	0.119 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.872 °C	30.0 °C	0.128 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.987 %HR	29.7 %HR	-0.287 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.914 %HR	46.5 %HR	-2.414 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.722 %HR	61.4 %HR	-3.322 %HR	± 3.0 %HR	± 1.9E+00 %HR